



ANALYTICAL REPORT

Lab Number:	L2010394
Client:	Act Global 410 South River St. Calhoun, GA 30701
ATTN:	Fred Gregg
Phone:	(706) 629-4774
Project Name:	ACT GLOBAL TURF YARN/BACK/PU
Project Number:	EPA 537M
Report Date:	04/02/20

The original project report/data package is held by Alpha Analytical. This report/data package is paginated and should be reproduced only in its entirety. Alpha Analytical holds no responsibility for results and/or data that are not consistent with the original.

Certifications & Approvals: MA (M-MA030), NH NELAP (2062), CT (PH-0141), DoD (L2474), FL (E87814), IL (200081), LA (85084), ME (MA00030), MD (350), NJ (MA015), NY (11627), NC (685), OH (CL106), PA (68-02089), RI (LAO00299), TX (T104704419), VT (VT-0015), VA (460194), WA (C954), US Army Corps of Engineers, USDA (Permit #P330-17-00150), USFWS (Permit #206964).

320 Forbes Boulevard, Mansfield, MA 02048-1806
508-822-9300 (Fax) 508-822-3288 800-624-9220 - www.alphalab.com



Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Alpha Sample ID	Client ID	Matrix	Sample Location	Collection Date/Time	Receive Date
L2010394-01	SAMPLE 1	SOLID	Not Specified		03/09/20

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Case Narrative

The samples were received in accordance with the Chain of Custody and no significant deviations were encountered during the preparation or analysis unless otherwise noted. Sample Receipt, Container Information, and the Chain of Custody are located at the back of the report.

Results contained within this report relate only to the samples submitted under this Alpha Lab Number and meet NELAP requirements for all NELAP accredited parameters unless otherwise noted in the following narrative. The data presented in this report is organized by parameter (i.e. VOC, SVOC, etc.). Sample specific Quality Control data (i.e. Surrogate Spike Recovery) is reported at the end of the target analyte list for each individual sample, followed by the Laboratory Batch Quality Control at the end of each parameter. Tentatively Identified Compounds (TICs), if requested, are reported for compounds identified to be present and are not part of the method/program Target Compound List, even if only a subset of the TCL are being reported. If a sample was re-analyzed or re-extracted due to a required quality control corrective action and if both sets of data are reported, the Laboratory ID of the re-analysis or re-extraction is designated with an "R" or "RE", respectively.

When multiple Batch Quality Control elements are reported (e.g. more than one LCS), the associated samples for each element are noted in the grey shaded header line of each data table. Any Laboratory Batch, Sample Specific % recovery or RPD value that is outside the listed Acceptance Criteria is bolded in the report. In reference to questions H (CAM) or 4 (RCP) when "NO" is checked, the performance criteria for CAM and RCP methods allow for some quality control failures to occur and still be within method compliance. In these instances, the specific failure is not narrated but noted in the associated QC Outlier Summary Report, located directly after the Case Narrative. QC information is also incorporated in the Data Usability Assessment table (Format 11) of our Data Merger tool, where it can be reviewed in conjunction with the sample result, associated regulatory criteria and any associated data usability implications.

Soil/sediments, solids and tissues are reported on a dry weight basis unless otherwise noted. Definitions of all data qualifiers and acronyms used in this report are provided in the Glossary located at the back of the report.

HOLD POLICY - For samples submitted on hold, Alpha's policy is to hold samples (with the exception of Air canisters) free of charge for 21 calendar days from the date the project is completed. After 21 calendar days, we will dispose of all samples submitted including those put on hold unless you have contacted your Alpha Project Manager and made arrangements for Alpha to continue to hold the samples. Air canisters will be disposed after 3 business days from the date the project is completed.

Please contact Project Management at 800-624-9220 with any questions.

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Case Narrative (continued)

Report Revision

April 2, 2020: A project name and number have been added.

March 27, 2020: The compound list has been revised.

Sample Receipt

The samples were received at the laboratory above the required temperature range and were not on ice.

Perfluorinated Alkyl Acids by Isotope Dilution

L2010394-01: The sample has elevated detection limits due to the limited sample volume utilized during extraction, as required by the sample matrix.

L2010394-01: Extracted Internal Standard recoveries were outside the acceptance criteria for individual analytes. Please refer to the surrogate section of the report for details.


The WG1353986-2 LCS recovery, associated with L2010394-01, is above the acceptance criteria for 1h,1h,2h,2h-perfluorodecanesulfonic acid (8:2fts) (161%); however, the associated sample is non-detect to the RL for this target analyte. The results of the original analysis are reported.

WG1353986-4 and WG1353986-5: Extracted Internal Standard recoveries were outside the acceptance criteria for individual analytes. Please refer to the surrogate section of the report for details.

WG1353986-5: The sample has elevated detection limits due to the limited sample volume utilized during extraction, as required by the sample matrix.

I, the undersigned, attest under the pains and penalties of perjury that, to the best of my knowledge and belief and based upon my personal inquiry of those responsible for providing the information contained in this analytical report, such information is accurate and complete. This certificate of analysis is not complete unless this page accompanies any and all pages of this report.

Authorized Signature:



Susan O'Neil

Title: Technical Director/Representative

Date: 04/02/20

ORGANICS

SEMIVOLATILES

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

SAMPLE RESULTS

Lab ID: L2010394-01
 Client ID: SAMPLE 1
 Sample Location: Not Specified

Date Collected:
 Date Received: 03/09/20
 Field Prep: Not Specified

Sample Depth:

Matrix: Solid
 Analytical Method: 134,LCMSMS-ID
 Analytical Date: 03/24/20 04:44
 Analyst: JW
 Percent Solids: Results reported on an 'AS RECEIVED' basis.

Extraction Method: ALPHA 23528
 Extraction Date: 03/23/20 09:15

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab						
Perfluorobutanoic Acid (PFBA)	ND		ng/g	1.67	--	1
Perfluoropentanoic Acid (PFPeA)	ND		ng/g	1.67	--	1
Perfluorobutanesulfonic Acid (PFBS)	ND		ng/g	1.67	--	1
Perfluorohexanoic Acid (PFHxA)	ND		ng/g	1.67	--	1
Perfluoroheptanoic Acid (PFHpA)	ND		ng/g	1.67	--	1
Perfluorohexanesulfonic Acid (PFHxS)	ND		ng/g	1.67	--	1
Perfluorooctanoic Acid (PFOA)	ND		ng/g	1.67	--	1
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND		ng/g	1.67	--	1
Perfluoroheptanesulfonic Acid (PFHpS)	ND		ng/g	1.67	--	1
Perfluorononanoic Acid (PFNA)	ND		ng/g	1.67	--	1
Perfluorooctanesulfonic Acid (PFOS)	ND		ng/g	1.67	--	1
Perfluorodecanoic Acid (PFDA)	ND		ng/g	1.67	--	1
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/g	1.67	--	1
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/g	1.67	--	1
Perfluoroundecanoic Acid (PFUnA)	ND		ng/g	1.67	--	1
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/g	1.67	--	1
Perfluorooctanesulfonamide (FOSA)	ND		ng/g	1.67	--	1
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND		ng/g	1.67	--	1
Perfluorododecanoic Acid (PFDoA)	ND		ng/g	1.67	--	1
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/g	1.67	--	1
Perfluorotetradecanoic Acid (PFTA)	ND		ng/g	1.67	--	1

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

SAMPLE RESULTS

Lab ID: L2010394-01
 Client ID: SAMPLE 1
 Sample Location: Not Specified

Date Collected:
 Date Received: 03/09/20
 Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab						
Surrogate (Extracted Internal Standard)			% Recovery	Qualifier		Acceptance Criteria
Perfluoro[13C4]Butanoic Acid (MPFBA)			76			60-153
Perfluoro[13C5]Pentanoic Acid (M5PFPEA)			91			65-182
Perfluoro[2,3,4-13C3]Butanesulfonic Acid (M3PFBS)			70			70-151
1H,1H,2H,2H-Perfluoro[1,2-13C2]Hexanesulfonic Acid (M2-4:2FTS)			81			56-138
Perfluoro[1,2,3,4,6-13C5]Hexanoic Acid (M5PFHxA)			75			61-147
Perfluoro[1,2,3,4-13C4]Heptanoic Acid (M4PFHpA)			75			62-149
Perfluoro[1,2,3-13C3]Hexanesulfonic Acid (M3PFHxS)			74			63-166
Perfluoro[13C8]Octanoic Acid (M8PFOA)			83			62-152
1H,1H,2H,2H-Perfluoro[1,2-13C2]Octanesulfonic Acid (M2-6:2FTS)			124			32-182
Perfluoro[13C9]Nonanoic Acid (M9PFNA)			87			61-154
Perfluoro[13C8]Octanesulfonic Acid (M8PFOS)			77			65-151
Perfluoro[1,2,3,4,5,6-13C6]Decanoic Acid (M6PFDA)			84			65-150
1H,1H,2H,2H-Perfluoro[1,2-13C2]Decanesulfonic Acid (M2-8:2FTS)			244	Q		25-186
N-Deuteriomethylperfluoro-1-octanesulfonamidoacetic Acid (d3-NMeFOSAA)			46			45-137
Perfluoro[1,2,3,4,5,6,7-13C7]Undecanoic Acid (M7-PFUDA)			69			64-158
Perfluoro[13C8]Octanesulfonamide (M8FOSA)			55			1-125
N-Deuterioethylperfluoro-1-octanesulfonamidoacetic Acid (d5-NEtFOSAA)			48			42-136
Perfluoro[1,2-13C2]Dodecanoic Acid (MPFDOA)			68			56-148
Perfluoro[1,2-13C2]Tetradecanoic Acid (M2PFTEDA)			70			26-160

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Method Blank Analysis
Batch Quality Control

Analytical Method: 134,LCMSMS-ID
Analytical Date: 03/24/20 04:11
Analyst: JW

Extraction Method: ALPHA 23528
Extraction Date: 03/23/20 09:15

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab for sample(s): 01 Batch: WG1353986-1					
Perfluorobutanoic Acid (PFBA)	ND		ng/g	1.00	--
Perfluoropentanoic Acid (PFPeA)	ND		ng/g	1.00	--
Perfluorobutanesulfonic Acid (PFBS)	ND		ng/g	1.00	--
Perfluorohexanoic Acid (PFHxA)	ND		ng/g	1.00	--
Perfluoroheptanoic Acid (PFHpA)	ND		ng/g	1.00	--
Perfluorohexanesulfonic Acid (PFHxS)	ND		ng/g	1.00	--
Perfluorooctanoic Acid (PFOA)	ND		ng/g	1.00	--
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND		ng/g	1.00	--
Perfluoroheptanesulfonic Acid (PFHpS)	ND		ng/g	1.00	--
Perfluorononanoic Acid (PFNA)	ND		ng/g	1.00	--
Perfluorooctanesulfonic Acid (PFOS)	ND		ng/g	1.00	--
Perfluorodecanoic Acid (PFDA)	ND		ng/g	1.00	--
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/g	1.00	--
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/g	1.00	--
Perfluoroundecanoic Acid (PFUnA)	ND		ng/g	1.00	--
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/g	1.00	--
Perfluorooctanesulfonamide (FOSA)	ND		ng/g	1.00	--
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND		ng/g	1.00	--
Perfluorododecanoic Acid (PFDoA)	ND		ng/g	1.00	--
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/g	1.00	--
Perfluorotetradecanoic Acid (PFTA)	ND		ng/g	1.00	--

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Method Blank Analysis
Batch Quality Control

Analytical Method: 134,LCMSMS-ID
Analytical Date: 03/24/20 04:11
Analyst: JW

Extraction Method: ALPHA 23528
Extraction Date: 03/23/20 09:15

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab for sample(s): 01 Batch: WG1353986-1					

Surrogate (Extracted Internal Standard)	%Recovery	Qualifier	Acceptance Criteria
Perfluoro[13C4]Butanoic Acid (MPFBA)	97		60-153
Perfluoro[13C5]Pentanoic Acid (M5PFPEA)	106		65-182
Perfluoro[2,3,4-13C3]Butanesulfonic Acid (M3PFBS)	95		70-151
1H,1H,2H,2H-Perfluoro[1,2-13C2]Hexanesulfonic Acid (M2-4:2FTS)	72		56-138
Perfluoro[1,2,3,4,6-13C5]Hexanoic Acid (M5PFHxA)	102		61-147
Perfluoro[1,2,3,4-13C4]Heptanoic Acid (M4PFHpA)	101		62-149
Perfluoro[1,2,3-13C3]Hexanesulfonic Acid (M3PFHxS)	105		63-166
Perfluoro[13C8]Octanoic Acid (M8PFOA)	99		62-152
1H,1H,2H,2H-Perfluoro[1,2-13C2]Octanesulfonic Acid (M2-6:2FTS)	66		32-182
Perfluoro[13C9]Nonanoic Acid (M9PFNA)	105		61-154
Perfluoro[13C8]Octanesulfonic Acid (M8PFOS)	103		65-151
Perfluoro[1,2,3,4,5,6-13C6]Decanoic Acid (M6PFDA)	103		65-150
1H,1H,2H,2H-Perfluoro[1,2-13C2]Decanesulfonic Acid (M2-8:2FTS)	88		25-186
N-Deuteriomethylperfluoro-1-octanesulfonamidoacetic Acid (d3-NMeFOSAA)	68		45-137
Perfluoro[1,2,3,4,5,6,7-13C7]Undecanoic Acid (M7-PFUDA)	101		64-158
Perfluoro[13C8]Octanesulfonamide (M8FOSA)	65		1-125
N-Deuterioethylperfluoro-1-octanesulfonamidoacetic Acid (d5-NEtFOSAA)	70		42-136
Perfluoro[1,2-13C2]Dodecanoic Acid (MPFDOA)	91		56-148
Perfluoro[1,2-13C2]Tetradecanoic Acid (M2PFTEDA)	88		26-160

Lab Control Sample Analysis

Batch Quality Control

Project Name: ACT GLOBAL TURF YARN/BACK/PU

Lab Number: L2010394

Project Number: EPA 537M

Report Date: 04/02/20

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab Associated sample(s): 01 Batch: WG1353986-2 WG1353986-3								
Perfluorobutanoic Acid (PFBA)	104		107		71-135	3		30
Perfluoropentanoic Acid (PFPeA)	111		113		69-132	2		30
Perfluorobutanesulfonic Acid (PFBS)	107		110		72-128	3		30
Perfluorohexanoic Acid (PFHxA)	103		105		70-132	2		30
Perfluoroheptanoic Acid (PFHpA)	106		108		71-131	2		30
Perfluorohexanesulfonic Acid (PFHxS)	114		107		67-130	6		30
Perfluorooctanoic Acid (PFOA)	106		111		69-133	5		30
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	114		130		64-140	13		30
Perfluoroheptanesulfonic Acid (PFHpS)	105		113		70-132	7		30
Perfluorononanoic Acid (PFNA)	106		109		72-129	3		30
Perfluorooctanesulfonic Acid (PFOS)	92		106		68-136	14		30
Perfluorodecanoic Acid (PFDA)	105		108		69-133	3		30
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	161	Q	130		65-137	21		30
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	111		114		63-144	3		30
Perfluoroundecanoic Acid (PFUnA)	109		108		64-136	1		30
Perfluorodecanesulfonic Acid (PFDS)	107		122		59-134	13		30
Perfluorooctanesulfonamide (FOSA)	111		108		67-137	3		30
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	110		107		61-139	3		30
Perfluorododecanoic Acid (PFDoA)	115		112		69-135	3		30
Perfluorotridecanoic Acid (PFTrDA)	126		129		66-139	2		30
Perfluorotetradecanoic Acid (PFTA)	110		117		69-133	6		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: ACT GLOBAL TURF YARN/BACK/PU

Lab Number: L2010394

Project Number: EPA 537M

Report Date: 04/02/20

Parameter	LCS		LCSD		%Recovery		RPD	RPD	
	%Recovery	Qual	%Recovery	Qual	Limits			Qual	Limits
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab Associated sample(s): 01 Batch: WG1353986-2 WG1353986-3									

Surrogate (Extracted Internal Standard)	LCS		LCSD		Acceptance Criteria
	%Recovery	Qual	%Recovery	Qual	
Perfluoro[13C4]Butanoic Acid (MPFBA)	103		99		60-153
Perfluoro[13C5]Pentanoic Acid (M5PFPEA)	111		106		65-182
Perfluoro[2,3,4-13C3]Butanesulfonic Acid (M3PFBS)	123		93		70-151
1H,1H,2H,2H-Perfluoro[1,2-13C2]Hexanesulfonic Acid (M2-4:2FTS)	99		72		56-138
Perfluoro[1,2,3,4,6-13C5]Hexanoic Acid (M5PFHxA)	112		107		61-147
Perfluoro[1,2,3,4-13C4]Heptanoic Acid (M4PFHpA)	109		104		62-149
Perfluoro[1,2,3-13C3]Hexanesulfonic Acid (M3PFHxS)	116		94		63-166
Perfluoro[13C8]Octanoic Acid (M8PFOA)	108		105		62-152
1H,1H,2H,2H-Perfluoro[1,2-13C2]Octanesulfonic Acid (M2-6:2FTS)	100		75		32-182
Perfluoro[13C9]Nonanoic Acid (M9PFNA)	112		106		61-154
Perfluoro[13C8]Octanesulfonic Acid (M8PFOS)	134		96		65-151
Perfluoro[1,2,3,4,5,6-13C6]Decanoic Acid (M6PFDA)	104		102		65-150
1H,1H,2H,2H-Perfluoro[1,2-13C2]Decanesulfonic Acid (M2-8:2FTS)	98		85		25-186
N-Deuteriomethylperfluoro-1-octanesulfonamidoacetic Acid (d3-NMeFOSAA)	77		83		45-137
Perfluoro[1,2,3,4,5,6,7-13C7]Undecanoic Acid (M7-PFUDA)	112		107		64-158
Perfluoro[13C8]Octanesulfonamide (M8FOSA)	66		70		1-125
N-Deuterioethylperfluoro-1-octanesulfonamidoacetic Acid (d5-NEtFOSAA)	82		78		42-136
Perfluoro[1,2-13C2]Dodecanoic Acid (MPFDOA)	102		95		56-148
Perfluoro[1,2-13C2]Tetradecanoic Acid (M2PFTEDA)	98		92		26-160

Matrix Spike Analysis

Batch Quality Control

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

<i>Parameter</i>	<i>Native Sample</i>	<i>MS Added</i>	<i>MS Found</i>	<i>MS %Recovery</i>	<i>Qual</i>	<i>MSD Found</i>	<i>MSD %Recovery</i>	<i>Qual</i>	<i>Recovery Limits</i>	<i>RPD</i>	<i>Qual</i>	<i>RPD Limits</i>
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab Associated sample(s): 01 QC Batch ID: WG1353986-4 QC Sample: L2010394-01 Client ID: SAMPLE 1												
Perfluorobutanoic Acid (PFBA)	ND	6.8	7.14	105		-	-		71-135	-		30
Perfluoropentanoic Acid (PFPeA)	ND	6.8	7.54	111		-	-		69-132	-		30
Perfluorobutanesulfonic Acid (PFBS)	ND	6.03	6.24	104		-	-		72-128	-		30
Perfluorohexanoic Acid (PFHxA)	ND	6.8	7.04	103		-	-		70-132	-		30
Perfluoroheptanoic Acid (PFHpA)	ND	6.8	7.15	105		-	-		71-131	-		30
Perfluorohexanesulfonic Acid (PFHxS)	ND	6.2	6.13	99		-	-		67-130	-		30
Perfluorooctanoic Acid (PFOA)	ND	6.8	7.44	109		-	-		69-133	-		30
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND	6.46	7.30	113		-	-		64-140	-		30
Perfluoroheptanesulfonic Acid (PFHpS)	ND	6.46	7.04	109		-	-		70-132	-		30
Perfluorononanoic Acid (PFNA)	ND	6.8	7.45	110		-	-		72-129	-		30
Perfluorooctanesulfonic Acid (PFOS)	ND	6.3	5.72	91		-	-		68-136	-		30
Perfluorodecanoic Acid (PFDA)	ND	6.8	7.13	105		-	-		69-133	-		30
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND	6.53	8.29	127		-	-		65-137	-		30
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND	6.8	8.02	118		-	-		63-144	-		30
Perfluoroundecanoic Acid (PFUnA)	ND	6.8	7.34	108		-	-		64-136	-		30
Perfluorodecanesulfonic Acid (PFDS)	ND	6.57	7.53	115		-	-		59-134	-		30
Perfluorooctanesulfonamide (FOSA)	ND	6.8	7.36	108		-	-		67-137	-		30
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND	6.8	7.82	115		-	-		61-139	-		30
Perfluorododecanoic Acid (PFDoA)	ND	6.8	7.92	116		-	-		69-135	-		30
Perfluorotridecanoic Acid (PFTrDA)	ND	6.8	8.69	128		-	-		66-139	-		30
Perfluorotetradecanoic Acid (PFTTA)	ND	6.8	7.83	115		-	-		69-133	-		30

Matrix Spike Analysis

Batch Quality Control

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Parameter	Native Sample	MS Added	MS Found	MS %Recovery	Qual	MSD Found	MSD %Recovery	Qual	Recovery Limits	RPD	Qual	RPD Limits
------------------	----------------------	-----------------	-----------------	---------------------	-------------	------------------	----------------------	-------------	------------------------	------------	-------------	-------------------

Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab Associated sample(s): 01 QC Batch ID: WG1353986-4 QC Sample: L2010394-01 Client ID: SAMPLE 1

Surrogate (Extracted Internal Standard)	MS % Recovery	MS Qualifier	MSD % Recovery	MSD Qualifier	Acceptance Criteria
1H,1H,2H,2H-Perfluoro[1,2-13C2]Decanesulfonic Acid (M2-8:2FTS)	291	Q			25-186
1H,1H,2H,2H-Perfluoro[1,2-13C2]Hexanesulfonic Acid (M2-4:2FTS)	109				56-138
1H,1H,2H,2H-Perfluoro[1,2-13C2]Octanesulfonic Acid (M2-6:2FTS)	169				32-182
N-Deuterioethylperfluoro-1-octanesulfonamidoacetic Acid (d5-NEtFOSAA)	59				42-136
N-Deuteriomethylperfluoro-1-octanesulfonamidoacetic Acid (d3-NMeFOSAA)	55				45-137
Perfluoro[1,2,3,4,5,6,7-13C7]Undecanoic Acid (M7-PFUDA)	81				64-158
Perfluoro[1,2,3,4,5,6-13C6]Decanoic Acid (M6PFDA)	97				65-150
Perfluoro[1,2,3,4,6-13C5]Hexanoic Acid (M5PFHxA)	90				61-147
Perfluoro[1,2,3,4-13C4]Heptanoic Acid (M4PFHpA)	89				62-149
Perfluoro[1,2,3-13C3]Hexanesulfonic Acid (M3PFHxS)	82				63-166
Perfluoro[1,2-13C2]Dodecanoic Acid (MPFDOA)	78				56-148
Perfluoro[1,2-13C2]Tetradecanoic Acid (M2PFTEDA)	86				26-160
Perfluoro[13C4]Butanoic Acid (MPFBA)	88				60-153
Perfluoro[13C5]Pentanoic Acid (M5PFPEA)	104				65-182
Perfluoro[13C8]Octanesulfonamide (M8FOSA)	62				1-125
Perfluoro[13C8]Octanesulfonic Acid (M8PFOS)	90				65-151
Perfluoro[13C8]Octanoic Acid (M8PFOA)	96				62-152
Perfluoro[13C9]Nonanoic Acid (M9PFNA)	101				61-154
Perfluoro[2,3,4-13C3]Butanesulfonic Acid (M3PFBS)	79				70-151

Lab Duplicate Analysis

Batch Quality Control

Project Name: ACT GLOBAL TURF YARN/BACK/PU

Project Number: EPA 537M

Lab Number: L2010394

Report Date: 04/02/20

Parameter	Native Sample	Duplicate Sample	Units	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab Associated sample(s): 01 QC Batch ID: WG1353986-5 QC Sample: L2010394-01 Client ID: SAMPLE 1						
Perfluorobutanoic Acid (PFBA)	ND	ND	ng/g	NC		30
Perfluoropentanoic Acid (PFPeA)	ND	ND	ng/g	NC		30
Perfluorobutanesulfonic Acid (PFBS)	ND	ND	ng/g	NC		30
Perfluorohexanoic Acid (PFHxA)	ND	ND	ng/g	NC		30
Perfluoroheptanoic Acid (PFHpA)	ND	ND	ng/g	NC		30
Perfluorohexanesulfonic Acid (PFHxS)	ND	ND	ng/g	NC		30
Perfluorooctanoic Acid (PFOA)	ND	ND	ng/g	NC		30
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND	ND	ng/g	NC		30
Perfluoroheptanesulfonic Acid (PFHpS)	ND	ND	ng/g	NC		30
Perfluorononanoic Acid (PFNA)	ND	ND	ng/g	NC		30
Perfluorooctanesulfonic Acid (PFOS)	ND	ND	ng/g	NC		30
Perfluorodecanoic Acid (PFDA)	ND	ND	ng/g	NC		30
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND	ND	ng/g	NC		30
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND	ND	ng/g	NC		30
Perfluoroundecanoic Acid (PFUnA)	ND	ND	ng/g	NC		30
Perfluorodecanesulfonic Acid (PFDS)	ND	ND	ng/g	NC		30
Perfluorooctanesulfonamide (FOSA)	ND	ND	ng/g	NC		30
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND	ND	ng/g	NC		30
Perfluorododecanoic Acid (PFDoA)	ND	ND	ng/g	NC		30
Perfluorotridecanoic Acid (PFTrDA)	ND	ND	ng/g	NC		30

Lab Duplicate Analysis

Batch Quality Control

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Parameter	Native Sample	Duplicate Sample	Units	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by Isotope Dilution - Mansfield Lab Associated sample(s): 01 QC Batch ID: WG1353986-5 QC Sample: L2010394-01 Client ID: SAMPLE 1						
Perfluorotetradecanoic Acid (PFTA)	ND	ND	ng/g	NC		30

Surrogate (Extracted Internal Standard)	%Recovery	Qualifier	%Recovery	Qualifier	Acceptance Criteria
Perfluoro[13C4]Butanoic Acid (MPFBA)	76		75		60-153
Perfluoro[13C5]Pentanoic Acid (M5PFPEA)	91		91		65-182
Perfluoro[2,3,4-13C3]Butanesulfonic Acid (M3PFBS)	70		68	Q	70-151
1H,1H,2H,2H-Perfluoro[1,2-13C2]Hexanesulfonic Acid (M2-4:2FTS)	81		95		56-138
Perfluoro[1,2,3,4,6-13C5]Hexanoic Acid (M5PFHxA)	75		76		61-147
Perfluoro[1,2,3,4-13C4]Heptanoic Acid (M4PFHpA)	75		76		62-149
Perfluoro[1,2,3-13C3]Hexanesulfonic Acid (M3PFHxS)	74		77		63-166
Perfluoro[13C8]Octanoic Acid (M8PFOA)	83		83		62-152
1H,1H,2H,2H-Perfluoro[1,2-13C2]Octanesulfonic Acid (M2-6:2FTS)	124		139		32-182
Perfluoro[13C9]Nonanoic Acid (M9PFNA)	87		90		61-154
Perfluoro[13C8]Octanesulfonic Acid (M8PFOS)	77		81		65-151
Perfluoro[1,2,3,4,5,6-13C6]Decanoic Acid (M6PFDA)	84		87		65-150
1H,1H,2H,2H-Perfluoro[1,2-13C2]Decanesulfonic Acid (M2-8:2FTS)	244	Q	297	Q	25-186
N-Deuteriomethylperfluoro-1-octanesulfonamidoacetic Acid (d3-NMeFOSAA)	46		48		45-137
Perfluoro[1,2,3,4,5,6,7-13C7]Undecanoic Acid (M7-PFUDA)	69		70		64-158
Perfluoro[13C8]Octanesulfonamide (M8FOSA)	55		62		1-125
N-Deuterioethylperfluoro-1-octanesulfonamidoacetic Acid (d5-NEtFOSAA)	48		51		42-136
Perfluoro[1,2-13C2]Dodecanoic Acid (MPFDOA)	68		67		56-148
Perfluoro[1,2-13C2]Tetradecanoic Acid (M2PFTEDA)	70		75		26-160

Project Name: ACT GLOBAL TURF YARN/BACK/PU

Project Number: EPA 537M

Sample Receipt and Container Information

Were project specific reporting limits specified?

YES

Cooler Information**Cooler** **Custody Seal**

A Absent

Container Information**Container ID** **Container Type**

L2010394-01A Bag

Cooler	Initial pH	Final pH	Temp deg C	Pres	Seal	Frozen Date/Time	Analysis(*)
A	NA		23.1	Y	Absent		A2-537-ISOTOPE(28)

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Serial_No:04022017:19
Lab Number: L2010394
Report Date: 04/02/20

PFAS PARAMETER SUMMARY

Parameter	Acronym	CAS Number
PERFLUOROALKYL CARBOXYLIC ACIDS (PFCAs)		
Perfluorooctadecanoic Acid	PFODA	16517-11-6
Perfluorohexadecanoic Acid	PFHxDA	67905-19-5
Perfluorotetradecanoic Acid	PFTA	376-06-7
Perfluorotridecanoic Acid	PFTrDA	72629-94-8
Perfluorododecanoic Acid	PFDoA	307-55-1
Perfluoroundecanoic Acid	PFUnA	2058-94-8
Perfluorodecanoic Acid	PFDA	335-76-2
Perfluorononanoic Acid	PFNA	375-95-1
Perfluorooctanoic Acid	PFOA	335-67-1
Perfluoroheptanoic Acid	PFHpA	375-85-9
Perfluorohexanoic Acid	PFHxA	307-24-4
Perfluoropentanoic Acid	PFPeA	2706-90-3
Perfluorobutanoic Acid	PFBA	375-22-4
PERFLUOROALKYL SULFONIC ACIDS (PFSAs)		
Perfluorododecanesulfonic Acid	PFDoDS	79780-39-5
Perfluorodecanesulfonic Acid	PFDS	335-77-3
Perfluoronanesulfonic Acid	PFNS	68259-12-1
Perfluorooctanesulfonic Acid	PFOS	1763-23-1
Perfluoroheptanesulfonic Acid	PFHpS	375-92-8
Perfluorohexanesulfonic Acid	PFHxS	355-46-4
Perfluoropentanesulfonic Acid	PFPeS	2706-91-4
Perfluorobutanesulfonic Acid	PFBS	375-73-5
FLUOROTELOMERS		
1H,1H,2H,2H-Perfluorododecanesulfonic Acid	10:2FTS	120226-60-0
1H,1H,2H,2H-Perfluorodecanesulfonic Acid	8:2FTS	39108-34-4
1H,1H,2H,2H-Perfluorooctanesulfonic Acid	6:2FTS	27619-97-2
1H,1H,2H,2H-Perfluorohexanesulfonic Acid	4:2FTS	757124-72-4
PERFLUOROALKANE SULFONAMIDES (FASAs)		
Perfluorooctanesulfonamide	FOSA	754-91-6
N-Ethyl Perfluorooctane Sulfonamide	NEtFOSA	4151-50-2
N-Methyl Perfluorooctane Sulfonamide	NMeFOSA	31506-32-8
PERFLUOROALKANE SULFONYL SUBSTANCES		
N-Ethyl Perfluorooctanesulfonamido Ethanol	NEtFOSE	1691-99-2
N-Methyl Perfluorooctanesulfonamido Ethanol	NMeFOSE	24448-09-7
N-Ethyl Perfluorooctanesulfonamidoacetic Acid	NEtFOSAA	2991-50-6
N-Methyl Perfluorooctanesulfonamidoacetic Acid	NMeFOSAA	2355-31-9
PER- and POLYFLUOROALKYL ETHER CARBOXYLIC ACIDS		
2,3,3,3-Tetrafluoro-2-[1,1,2,2,3,3,3-Heptafluoropropoxy]-Propanoic Acid	HFPO-DA	13252-13-6
4,8-Dioxa-3h-Perfluorononanoic Acid	ADONA	919005-14-4
CHLORO-PERFLUOROALKYL SULFONIC ACIDS		
11-Chloroicosafuoro-3-Oxaundecane-1-Sulfonic Acid	11Cl-PF3OUdS	763051-92-9
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid	9Cl-PF3ONS	756426-58-1

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

GLOSSARY

Acronyms

DL	- Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the limit of quantitation (LOQ). The DL includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.)
EDL	- Estimated Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the reporting limit (RL). The EDL includes any adjustments from dilutions, concentrations or moisture content, where applicable. The use of EDLs is specific to the analysis of PAHs using Solid-Phase Microextraction (SPME).
EMPC	- Estimated Maximum Possible Concentration: The concentration that results from the signal present at the retention time of an analyte when the ions meet all of the identification criteria except the ion abundance ratio criteria. An EMPC is a worst-case estimate of the concentration.
EPA	- Environmental Protection Agency.
LCS	- Laboratory Control Sample: A sample matrix, free from the analytes of interest, spiked with verified known amounts of analytes or a material containing known and verified amounts of analytes.
LCSD	- Laboratory Control Sample Duplicate: Refer to LCS.
LFB	- Laboratory Fortified Blank: A sample matrix, free from the analytes of interest, spiked with verified known amounts of analytes or a material containing known and verified amounts of analytes.
LOD	- Limit of Detection: This value represents the level to which a target analyte can reliably be detected for a specific analyte in a specific matrix by a specific method. The LOD includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.)
LOQ	- Limit of Quantitation: The value at which an instrument can accurately measure an analyte at a specific concentration. The LOQ includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.) Limit of Quantitation: The value at which an instrument can accurately measure an analyte at a specific concentration. The LOQ includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.)
MDL	- Method Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the reporting limit (RL). The MDL includes any adjustments from dilutions, concentrations or moisture content, where applicable.
MS	- Matrix Spike Sample: A sample prepared by adding a known mass of target analyte to a specified amount of matrix sample for which an independent estimate of target analyte concentration is available. For Method 332.0, the spike recovery is calculated using the native concentration, including estimated values.
MSD	- Matrix Spike Sample Duplicate: Refer to MS.
NA	- Not Applicable.
NC	- Not Calculated: Term is utilized when one or more of the results utilized in the calculation are non-detect at the parameter's reporting unit.
NDPA/DPA	- N-Nitrosodiphenylamine/Diphenylamine.
NI	- Not Ignitable.
NP	- Non-Plastic: Term is utilized for the analysis of Atterberg Limits in soil.
RL	- Reporting Limit: The value at which an instrument can accurately measure an analyte at a specific concentration. The RL includes any adjustments from dilutions, concentrations or moisture content, where applicable.
RPD	- Relative Percent Difference: The results from matrix and/or matrix spike duplicates are primarily designed to assess the precision of analytical results in a given matrix and are expressed as relative percent difference (RPD). Values which are less than five times the reporting limit for any individual parameter are evaluated by utilizing the absolute difference between the values; although the RPD value will be provided in the report.
SRM	- Standard Reference Material: A reference sample of a known or certified value that is of the same or similar matrix as the associated field samples.
STLP	- Semi-dynamic Tank Leaching Procedure per EPA Method 1315.
TEF	- Toxic Equivalency Factors: The values assigned to each dioxin and furan to evaluate their toxicity relative to 2,3,7,8-TCDD.
TEQ	- Toxic Equivalent: The measure of a sample's toxicity derived by multiplying each dioxin and furan by its corresponding TEF and then summing the resulting values.
TIC	- Tentatively Identified Compound: A compound that has been identified to be present and is not part of the target compound list (TCL) for the method and/or program. All TICs are qualitatively identified and reported as estimated concentrations.

Footnotes

Report Format: Data Usability Report



Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

- 1 - The reference for this analyte should be considered modified since this analyte is absent from the target analyte list of the original method.

Terms

Analytical Method: Both the document from which the method originates and the analytical reference method. (Example: EPA 8260B is shown as 1,8260B.) The codes for the reference method documents are provided in the References section of the Addendum.

Difference: With respect to Total Oxidizable Precursor (TOP) Assay analysis, the difference is defined as the Post-Treatment value minus the Pre-Treatment value.

Final pH: As it pertains to Sample Receipt & Container Information section of the report, Final pH reflects pH of container determined after adjustment at the laboratory, if applicable. If no adjustment required, value reflects Initial pH.

Frozen Date/Time: With respect to Volatile Organics in soil, Frozen Date/Time reflects the date/time at which associated Reagent Water-preserved vials were initially frozen. Note: If frozen date/time is beyond 48 hours from sample collection, value will be reflected in 'bold'.

Initial pH: As it pertains to Sample Receipt & Container Information section of the report, Initial pH reflects pH of container determined upon receipt, if applicable.

PAH Total: With respect to Alkylated PAH analyses, the 'PAHs, Total' result is defined as the summation of results for all or a subset of the following compounds: Naphthalene, C1-C4 Naphthalenes, 2-Methylnaphthalene, 1-Methylnaphthalene, Biphenyl, Acenaphthylene, Acenaphthene, Fluorene, C1-C3 Fluorenes, Phenanthrene, C1-C4 Phenanthrenes/Anthracenes, Anthracene, Fluoranthene, Pyrene, C1-C4 Fluoranthenes/Pyrenes, Benz(a)anthracene, Chrysene, C1-C4 Chrysenes, Benzo(b)fluoranthene, Benzo(j)+(k)fluoranthene, Benzo(e)pyrene, Benzo(a)pyrene, Perylene, Indeno(1,2,3-cd)pyrene, Dibenz(ah)+(ac)anthracene, Benzo(g,h,i)perylene. If a 'Total' result is requested, the results of its individual components will also be reported.

PFAS Total: With respect to PFAS analyses, the 'PFAS, Total (5)' result is defined as the summation of results for: PFHpA, PFHxS, PFOA, PFNA and PFOS. If a 'Total' result is requested, the results of its individual components will also be reported.

The target compound Chlordane (CAS No. 57-74-9) is reported for GC ECD analyses. Per EPA, this compound "refers to a mixture of chlordane isomers, other chlorinated hydrocarbons and numerous other components." (Reference: USEPA Toxicological Review of Chlordane, In Support of Summary Information on the Integrated Risk Information System (IRIS), December 1997.)

Total: With respect to Organic analyses, a 'Total' result is defined as the summation of results for individual isomers or Aroclors. If a 'Total' result is requested, the results of its individual components will also be reported. This is applicable to 'Total' results for methods 8260, 8081 and 8082.

Data Qualifiers

- A** - Spectra identified as "Aldol Condensates" are byproducts of the extraction/concentration procedures when acetone is introduced in the process.
- B** - The analyte was detected above the reporting limit in the associated method blank. Flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank. For MCP-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank. For DOD-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank AND the analyte was detected above one-half the reporting limit (or above the reporting limit for common lab contaminants) in the associated method blank. For NJ-Air-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte above the reporting limit. For NJ-related projects (excluding Air), flag only applies to associated field samples that have detectable concentrations of the analyte, which was detected above the reporting limit in the associated method blank or above five times the reporting limit for common lab contaminants (Phthalates, Acetone, Methylene Chloride, 2-Butanone).
- C** - Co-elution: The target analyte co-elutes with a known lab standard (i.e. surrogate, internal standards, etc.) for co-extracted analyses.
- D** - Concentration of analyte was quantified from diluted analysis. Flag only applies to field samples that have detectable concentrations of the analyte.
- E** - Concentration of analyte exceeds the range of the calibration curve and/or linear range of the instrument.
- G** - The concentration may be biased high due to matrix interferences (i.e. co-elution) with non-target compound(s). The result should be considered estimated.
- H** - The analysis of pH was performed beyond the regulatory-required holding time of 15 minutes from the time of sample collection.
- I** - The lower value for the two columns has been reported due to obvious interference.
- J** - Estimated value. This represents an estimated concentration for Tentatively Identified Compounds (TICs).
- M** - Reporting Limit (RL) exceeds the MCP CAM Reporting Limit for this analyte.
- ND** - Not detected at the reporting limit (RL) for the sample.
- NJ** - Presumptive evidence of compound. This represents an estimated concentration for Tentatively Identified Compounds (TICs), where the identification is based on a mass spectral library search.
- P** - The RPD between the results for the two columns exceeds the method-specified criteria.
- Q** - The quality control sample exceeds the associated acceptance criteria. For DOD-related projects, LCS and/or Continuing Calibration Standard exceedences are also qualified on all associated sample results. Note: This flag is not applicable for matrix spike recoveries when the sample concentration is greater than 4x the spike added or for batch duplicate RPD when the sample concentrations are less

Report Format: Data Usability Report



Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

Data Qualifiers

than 5x the RL. (Metals only.)

- R** - Analytical results are from sample re-analysis.
- RE** - Analytical results are from sample re-extraction.
- S** - Analytical results are from modified screening analysis.

Project Name: ACT GLOBAL TURF YARN/BACK/PU
Project Number: EPA 537M

Lab Number: L2010394
Report Date: 04/02/20

REFERENCES

- 134 Determination of Selected Perfluorinated Alkyl Acids in Drinking Water by Solid Phase Extraction and Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS) using Isotope Dilution. Alpha SOP 23528.

LIMITATION OF LIABILITIES

Alpha Analytical performs services with reasonable care and diligence normal to the analytical testing laboratory industry. In the event of an error, the sole and exclusive responsibility of Alpha Analytical shall be to re-perform the work at it's own expense. In no event shall Alpha Analytical be held liable for any incidental, consequential or special damages, including but not limited to, damages in any way connected with the use of, interpretation of, information or analysis provided by Alpha Analytical.

We strongly urge our clients to comply with EPA protocol regarding sample volume, preservation, cooling, containers, sampling procedures, holding time and splitting of samples in the field.



Certification Information

The following analytes are not included in our Primary NELAP Scope of Accreditation:

Westborough Facility

EPA 624/624.1: m/p-xylene, o-xylene

EPA 8260C: NPW: 1,2,4,5-Tetramethylbenzene; 4-Ethyltoluene, Azobenzene; SCM: Iodomethane (methyl iodide), 1,2,4,5-Tetramethylbenzene; 4-Ethyltoluene.

EPA 8270D: NPW: Dimethylnaphthalene, 1,4-Diphenylhydrazine; SCM: Dimethylnaphthalene, 1,4-Diphenylhydrazine.

SM4500: NPW: Amenable Cyanide; SCM: Total Phosphorus, TKN, NO₂, NO₃.

Mansfield Facility

SM 2540D: TSS

EPA 8082A: NPW: PCB: 1, 5, 31, 87, 101, 110, 141, 151, 153, 180, 183, 187.

EPA TO-15: Halothane, 2,4,4-Trimethyl-2-pentene, 2,4,4-Trimethyl-1-pentene, Thiophene, 2-Methylthiophene, 3-Methylthiophene, 2-Ethylthiophene, 1,2,3-Trimethylbenzene, Indan, Indene, 1,2,4,5-Tetramethylbenzene, Benzothiophene, 1-Methylnaphthalene.

EPA TO-12 Non-methane organics

EPA 3C Fixed gases

Biological Tissue Matrix: EPA 3050B

The following analytes are included in our Massachusetts DEP Scope of Accreditation

Westborough Facility:

Drinking Water

EPA 300.0: Chloride, Nitrate-N, Fluoride, Sulfate; **EPA 353.2:** Nitrate-N, Nitrite-N; **SM4500NO3-F:** Nitrate-N, Nitrite-N; **SM4500F-C, SM4500CN-CE,**

EPA 180.1, SM2130B, SM4500CI-D, SM2320B, SM2540C, SM4500H-B, SM4500NO2-B

EPA 332: Perchlorate; **EPA 524.2:** THMs and VOCs; **EPA 504.1:** EDB, DBCP.

Microbiology: **SM9215B; SM9223-P/A, SM9223B-Colilert-QT, SM9222D.**

Non-Potable Water

SM4500H,B, EPA 120.1, SM2510B, SM2540C, SM2320B, SM4500CL-E, SM4500F-BC, SM4500NH3-BH: Ammonia-N and Kjeldahl-N, **EPA 350.1:** Ammonia-N, **LACHAT 10-107-06-1-B:** Ammonia-N, **EPA 351.1, SM4500NO3-F, EPA 353.2:** Nitrate-N, **SM4500P-E, SM4500P-B, E, SM4500SO4-E, SM5220D, EPA 410.4, SM5210B, SM5310C, SM4500CL-D, EPA 1664, EPA 420.1, SM4500-CN-CE, SM2540D, EPA 300:** Chloride, Sulfate, Nitrate.

EPA 624.1: Volatile Halocarbons & Aromatics,

EPA 608.3: Chlordane, Toxaphene, Aldrin, alpha-BHC, beta-BHC, gamma-BHC, delta-BHC, Dieldrin, DDD, DDE, DDT, Endosulfan I, Endosulfan II, Endosulfan sulfate, Endrin, Endrin Aldehyde, Heptachlor, Heptachlor Epoxide, PCBs

EPA 625.1: SVOC (Acid/Base/Neutral Extractables), **EPA 600/4-81-045:** PCB-Oil.

Microbiology: **SM9223B-Colilert-QT; Enterolert-QT, SM9221E, EPA 1600, EPA 1603.**

Mansfield Facility:

Drinking Water

EPA 200.7: Al, Ba, Cd, Cr, Cu, Fe, Mn, Ni, Na, Ag, Ca, Zn. **EPA 200.8:** Al, Sb, As, Ba, Be, Cd, Cr, Cu, Pb, Mn, Ni, Se, Ag, TL, Zn. **EPA 245.1** Hg.

EPA 522.

Non-Potable Water

EPA 200.7: Al, Sb, As, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Mo, Ni, K, Se, Ag, Na, Sr, TL, Ti, V, Zn.

EPA 200.8: Al, Sb, As, Be, Cd, Cr, Cu, Fe, Pb, Mn, Ni, K, Se, Ag, Na, TL, Zn.

EPA 245.1 Hg.

SM2340B

For a complete listing of analytes and methods, please contact your Alpha Project Manager.

L2010394

To whom it may concern,

I am needing PFAS testing done on this sample. Please call me to discuss.

Thank you,

Fred Gregg



FRED GREGG
Quality Manager
fgregg@actglobal.com

T +1.706.629.4774
M +1.706.263.1400
410 South River St
Calhoun, GA 30701



ActGlobal.com

*Rec: 3/9/20 8:58am
Kimberly - AAC*